

Product Name:	ABS-DDL2-18
	<p> USB Digital Data Logger (DDL), complied to the current CDC vaccine storage monitor guidelines 3 years certification of calibration “buffered” probe in the product simulated solution °F/°C Switchable Readout Min/Max historical temp point Detachable Dual Probe High and Low Temp Alarm Visual Alarm Audible Alarm Operation Temp Range: -50.00 to 60.00°C (-58 to 140°F) Accuracy of +/-0 .25° C Readout resolution: 0.01 Low-Battery Indicator Temp Data downloaded via USB drive (not included) User programmable Logging Interval (recommend no less than every 30 mins) Store up to 3 years of data (30-minute intervals) </p>
Description:	